

| Channels | Part No. | Volume Range (µl) | Volume Increments (µl) | Test Volume (µl) | Accuracy (±%) | Precision (≤%) |
|----------|----------|-------------------|------------------------|------------------|---------------|----------------|
| 24 | 6123 | 10 – 300 | 0.5 | 30 | 2.50 | 1.50 |
| | | | | 150 | 1.50 | 1.00 |
| | | | | 300 | 1.20 | 0.80 |
| | 6124 | 50 – 1250 | 1 | 125 | 2.50 | 1.50 |
| | | | | 625 | 2.00 | 1.00 |
| | | | | 1250 | 1.80 | 0.80 |
| 96 | 6101 | 0.5 – 12.5 | 0.01 | 1.25 | 8.00 | 3.00 |
| | | | | 6.25 | 2.00 | 1.50 |
| | | | | 12.5 | 1.50 | 1.00 |
| | 6106 | 2 – 50 | 0.05 | 5 | 5.00 | 2.50 |
| | | | | 25 | 2.00 | 1.50 |
| | | | | 50 | 1.50 | 1.00 |
| | 6102 | 5 – 125 | 0.1 | 12.5 | 3.00 | 1.70 |
| | | | | 62.5 | 2.00 | 1.50 |
| | | | | 125 | 1.50 | 1.00 |
| | 6103 | 10 – 300 | 0.5 | 30 | 2.50 | 1.50 |
| | | | | 150 | 1.50 | 1.00 |
| | | | | 300 | 1.20 | 0.80 |
| | 6104 | 50 – 1250 | 1 | 125 | 2.50 | 1.50 |
| | | | | 625 | 2.00 | 1.00 |
| | | | | 1250 | 1.80 | 0.80 |
| 384 | 6131 | 0.5 – 12.5 | 0.01 | 1.25 | 8.00 | 3.00 |
| | | | | 6.25 | 2.00 | 1.50 |
| | | | | 12.5 | 1.50 | 1.00 |
| | 6136 | 2 – 50 | 0.05 | 5 | 5.00 | 2.50 |
| | | | | 25 | 2.00 | 1.50 |
| | | | | 50 | 1.50 | 1.00 |
| | 6132 | 5 – 125 | 0.1 | 12.5 | 3.00 | 1.70 |
| | | | | 62.5 | 2.00 | 1.50 |
| | | | | 125 | 1.50 | 1.00 |

1. Specifications represent the performance of all channels.
2. Specifications apply to neat transfers in Pipet mode only.
3. Precision = coefficient of variation