

Kanal	Art.-Nr.	Volumen-Bereich (µl)	Volumen-schritt (µl)	Test-Volumen (µl)	Genauigkeit (±%)	Präzision (≤%)	Spitzen-Abstand (mm)
4	4743	10 – 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 – 33.0
	4744	50 – 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 – 33.0
6	4763	10 – 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 – 19.8
	4764	50 – 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 – 19.8
8	4721	0.5 – 12.5	0.01	1.25 6.25 12.5	10.00 4.00 2.00	6.00 1.60 0.80	4.5 – 14.0
	4726	2 – 50	0.05	5 25 50	5.00 2.50 1.50	2.50 0.60 0.40	4.5 – 14.0
	4722	5 – 125	0.1	12.5 62.5 125	3.75 2.50 1.60	1.50 0.70 0.35	4.5 – 14.0
	4723	10 – 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 – 14.0
	4724	50 – 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 – 14.0
12	4731	0.5 – 12.5	0.01	1.25 6.25 12.5	10.00 4.00 2.00	6.00 1.60 0.80	4.5 – 9.0
	4736	2 – 50	0.05	5 25 50	5.00 2.50 1.50	2.50 0.60 0.40	4.5 – 9.0
	4732	5 – 125	0.1	12.5 62.5 125	3.75 2.50 1.60	1.50 0.70 0.35	4.5 – 9.0

1. Die Spezifikationen der Mehrkanal-Pipetten bilden die Eigenschaften aller Pipettierkanäle ab.

2. Präzision = Streukoeffizient